
From: Nathan Moskowitz <nmosk@hotmail.com>
Sent: Saturday, May 28, 2011 5:20 AM CDT
To: eric sugalski <eric@bostondevice.com>
Subject: RE: updated e-drawing

Eric,
Looks very nice.

The ridges are very important for bone integration, and the inserts are important for tool application. I will discuss with Andy over the weekend. I would tend to doubt that would affect fabrication costs.
Thanks,
Nathan

From: eric@bostondevice.com
Date: Fri, 27 May 2011 14:43:09 -0400
Subject: updated e-drawing
To: nmosk@hotmail.com; andymoskow@gmail.com
CC: dschoon@bostondevice.com

Hi Nathan & Andy -

Please review the attached e-drawing. There are still some interferences and fabrication details that need to get resolved, but we're looking for your feedback on the overall proportions / geometry before we finalize the design. Please let us know your thoughts.
Thanks,

Eric

--

Eric Sugalski
Boston Device Development, Inc
office - (617) 279-0057 x703
mobile - (617) 852-6622
www.bostondevice.com